



NABL

National Accreditation Board for Testing and Calibration Laboratories

Department of Science & Technology, India

CERTIFICATE OF ACCREDITATION

EMI SOLUTIONS PRIVATE LIMITED

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2005

"General Requirements for the Competence of Testing & Calibration Laboratories"

for its facilities at

No. 237, A-5, Bommasandra Industrial Area, Bangalore

in the discipline of

ELECTRONICS TESTING

(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Certificate Number T-0472

Issue Date 31/03/2012



Valid Until 30/03/2014

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the additional requirements of NABL.

Signed for and on behalf of NABL

Mallika

Mallika Gope
Convenor

Anil Relia

Anil Relia
Director

Thannan

Dr. T. Ramasami
Chairman



NABL

Department of Science & Technology, India

SCOPE OF ACCREDITATION

Laboratory	EMI Solutions Private Limited, No. 237, A-5, Bommasandra Industrial Area, Bangalore		
Accreditation Standard	ISO/IEC 17025: 2005		
Discipline	Electronics Testing	Issue Date	31.03.2012
Certificate Number	T-0472	Valid Until	30.03.2014
Last Amended on	-	Page	1 of 1

S.No.	Product / Material of Test	Specific Test Performed	Test Method Specification against which tests are performed	Range of Testing / Limits of Detection
1.	ELECTRONICS/ ELECTRICAL EQUIPMENTS	Conducted Emission	CISPR 11: 2010-05, Edition 5.1 CISPR 14-1: 2009-02, Edition 5.1 CISPR 14-2: 2008-07, Edition 1.2 CISPR 22: 2008-09, Edition 6.0	150kHz to 30MHz
		Electro Static Discharge (ESD)	IEC 61000-4-2: 2008-12 Edition 2.0	0.5 to 15 kV
		Electrical Fast Transients (Burst)	IEC 61000-4-4: 2011-03 Edition 2.1	0.25 to 4 kV on power Lines
		Surge	IEC 61000-4-5: 2005-11 Edition 2.0	0.5 to 4 kV on power Lines
		Line Voltage dips & Short Interruption	IEC 61000-4-11: 2004-03 Edition 2.0	230 V 0 to 80% dip ½ Cycle to 250 Cycles

-X-X-X-X-X-X-X-X-X-X-X-

mali'ka

Convenor